

File :Q:\GCMS-08 Minimal\DATA\20190925\15- .D
Operator : ZA / KM / JP
Acquired : 25 Sep 2019 17:21 using AcqMethod KERLA-2017-100uL draw speed.M
Instrument : GCMS 8
Sample Name: P-1007S
Misc Info : 1-Naphthol in MeOH
Vial Number: 3



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

