



=====

Acq. Operator : SYSTEM	Seq. Line : 6
Acq. Instrument : LCMS-1	Location : P1-B2
Injection Date : 7/13/2018 5:19:59 PM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : D:\DATA\07131801\07101801 2018-07-13 15-46-15\Purity195_315.M
Last changed : 7/13/2018 3:46:15 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 6/6/2017 9:45:25 AM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram (s)

