



=====
Acq. Operator : SYSTEM Seq. Line : 16

Acq. Instrument : HPLC only Location : P1-B6

Injection Date : 6/11/2019 3:19:18 PM Inj : 1

Inj Volume : 1.000 µl

Acq. Method : C:\Chem32\2\Data\06111901\06111901 2019-06-11 10-38-49\purity new.M

Last changed : 6/11/2019 10:38:49 AM by SYSTEM

Analysis Method : C:\CHEM32\2\METHODS\purity new.M

Last changed : 6/11/2019 11:41:27 AM by SYSTEM

(modified after loading)

Method Info : Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size

Sample Info : Ametoctradin

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

