Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\01082101\14-7-GA.D
Sample Name : P-1062S

==========

Acq. Operator : SYSTEM Seq. Line : 14
Acq. Instrument : LCMS-1 Location : P2-D1
Injection Date : 1/8/2021 4:19:18 PM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/8/2020 11:20:44 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 1/11/2021 11:03:12 AM by SYSTEM
(modified after loading)
Method Info : Operator: Oraganic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Current Chromatogram(s)