



=====

Acq. Operator	: SYSTEM	Seq. Line	: 16
Acq. Instrument	: LCMS-1	Location	: P1-D3
Injection Date	: 12/7/2022 11:07:59 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/9/2022 8:35:37 AM by SYSTEM

(modified after loading)

Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate : 0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple MS Detector. No UV Reference.

Sample Info : Spinetoram in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

