



=====

Acq. Operator : SYSTEM	Seq. Line : 5
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P1-B-04
Injection Date : 7/17/2020 6:52:35 AM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\07172001\07132001 2020-07-17 05-35-58\Purity_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µ particlue size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

