



=====

Acq. Operator	: SYSTEM	Seq. Line	: 26
Acq. Instrument	: LCMS-1	Location	: P1-C4
Injection Date	: 7/17/2018 7:57:48 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : D:\DATA\07171801\07171801 2018-07-17 12-11-55\Purity195_315.M
Last changed : 7/17/2018 12:11:55 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 6/6/2017 9:45:25 AM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

