

Print of window 38: Current Chromatogram(s)

Data File : D:\DATA\06121801\06121801 2018-06-12 07-01-45\16-

Sample Name : P-1099S-CN-10X (Benzobicyclon)



=====

Acq. Operator : SYSTEM	Seq. Line : 22
Acq. Instrument : LCMS-1	Location : P1-C1
Injection Date : 6/12/2018 1:46:54 PM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : D:\DATA\06121801\06121801 2018-06-12 07-01-45\Purity195\_315.M

Last changed : 6/12/2018 7:01:45 AM by SYSTEM

Analysis Method : C:\Chem32\1\Methods\Purity195\_315.M

Last changed : 6/6/2017 9:45:25 AM by SYSTEM

Method Info : Operator: GA

Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

