Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\04192021\08- -KD.D
Sample Name : P-1157S-CN-10X

AccuStandard®
Leader in Analytical Reference Standards

==========

Acq. Operator : SYSTEM Seq. Line : 8
Acq. Instrument : LCMS-1 Location : P1-D5
Injection Date : 4/19/2021 10:32:46 AM Inj : 1
Inj Volume : 1.000 µl
Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/8/2020 11:20:44 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 4/15/2021 1:39:32 PM by SYSTEM
(modified after loading)
Method Info : Operator: Oraganic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)