

```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    4
Acq. Instrument : LCMS-1                             Location  : P1-A3
Injection Date  : 6/16/2021 8:50:43 AM                Inj       :    1
                                                Inj Volume: 1.000 µl
```

```
Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 5/21/2021 4:03:06 PM by SYSTEM
Analysis Method: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 6/16/2021 10:33:12 AM by SYSTEM
                (modified after loading)
```

```
Method Info    : Operator: Organic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                MS Detector. No UV Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

