

=====
 Acq. Operator : SYSTEM Seq. Line : 4
 Sample Operator : SYSTEM
 Acq. Instrument : LCMS-2 Location : P2-A-06
 Injection Date : 6/29/2021 6:58:35 AM Inj : 1
 Inj Volume : 1.000 µl
 Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06292101\06292101 2021-06-29
 06-01-05\Purity_195-315.M
 Last changed : 4/29/2021 10:41:07 AM by SYSTEM
 Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
 Last changed : 4/29/2021 10:41:07 AM by SYSTEM
 Method Info : OPERATOR: Organic QC lab
 General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
 Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
 Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity
 II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

