

=====

Acq. Operator : SYSTEM	Seq. Line : 2
Acq. Instrument : LCMS-1	Location : P2-F5
Injection Date : 6/30/2022 12:06:07 PM	Inj : 2
	Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 11/5/2021 2:37:17 PM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 6/21/2022 4:09:45 PM by SYSTEM  
(modified after loading)

Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Tetrachlorvinphos in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

