

Print of window 38: Current Chromatogram(s)

Data File : D:\DATA\06261801\06261801 2018-06-26 12-27-44\

Sample Name : P-1365S-CN-10X Imidacloprid-urea



=====

| | | | |
|-----------------|------------------------|------------|------------|
| Acq. Operator | : SYSTEM | Seq. Line | : 12 |
| Acq. Instrument | : LCMS-1 | Location | : P2-F6 |
| Injection Date | : 6/26/2018 3:54:10 PM | Inj | : 1 |
| | | Inj Volume | : 1.000 µl |

Acq. Method : D:\DATA\06261801\06261801 2018-06-26 12-27-44\Purity195_315.M
Last changed : 6/26/2018 12:27:44 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 6/26/2018 11:59:55 AM by SYSTEM
(modified after loading)

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.
3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule
size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector.
No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

