

=====

Acq. Operator	: SYSTEM	Seq. Line	: 9
Sample Operator	: SYSTEM		
Acq. Instrument	: LCMS-2	Location	: P2-B-04
Injection Date	: 1/31/2020 1:19:06 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\01312001\01202001 2020-01-31 10-47-04
\Purity_195-315.M

Last changed : 12/3/2019 1:25:24 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 2/3/2020 10:49:33 AM by SYSTEM

(modified after loading)

Method Info : General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

