

File :D:\MassHunter\GCMS\1\DATA\03242301\02- 2ALR.D
Operator : Organic QC department
Acquired : 24 Mar 2023 07:28 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: P-148S
Misc Info : Fenthion in MeOH
Vial Number: 144



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

