



=====

Acq. Operator : SYSTEM	Seq. Line : 16
Acq. Instrument : LCMS-1	Location : P1-D1
Injection Date : 5/17/2023 11:17:06 AM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/16/2023 8:46:04 AM by SYSTEM
(modified after loading)

Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Sample Info : Terbutylazine in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

