

File :D:\MassHunter\GCMS\1\DATA\083121\02-ALR.D  
Operator : Organic QC Lab  
Acquired : 31 Aug 2021 09:46 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name : P-180S-10X  
Misc Info : Thiobencarb in MeOH  
Vial Number: 118



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

