

File :D:\MassHunter\GCMS\1\data\07122022\02 - SAK.D
Operator :
Acquired : 12 Jul 2022 08:21 using AcqMethod ALR 19 min 20 split FAST .M
Instrument : GCMS 9
Sample Name: P-191S-10X
Misc Info : Butachlor in MeOH
Vial Number: 2



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

