

File :D:\MassHunter\GCMS\1\data\01192024\34- ALR.D  
Operator :  
Acquired : 19 Jan 2024 18:55 using AcqMethod ALR 19 min 20 split FAST .M  
Instrument : GCMS 9  
Sample Name: P-197S-10X  
Misc Info : Trifluralin in MeOH  
Vial Number: 1



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

