

File :Q:\GCMS-08 Minimal\DATA\20210507\06 YX.D  
Operator : Organic QC Lab  
Acquired : 07 May 2021 11:55 using AcqMethod Hy\_PAH\_15cm\_S5.M  
Instrument : GCMS 8  
Sample Name: P-208S  
Misc Info : Terbufos in MeOH  
Vial Number: 134



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

