



```
=====  
Acq. Operator   : SYSTEM                      Seq. Line :    4  
Acq. Instrument : LCMS-1                      Location  : P1-C7  
Injection Date  : 6/2/2021 3:05:02 PM          Inj       :    1  
                                                    Inj Volume: 1.000 µl  
Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M  
Last changed    : 5/21/2021 4:03:06 PM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M  
Last changed    : 6/3/2021 2:31:55 PM by SYSTEM  
                (modified after loading)  
Method Info     : Operator: Organic QC lab  
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate  
                :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron  
                particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole  
                MS Detector. No UV Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

