

File :D:\MassHunter\GCMS\1\DATA\07072101\26- ALR.D  
Operator : Organic QC department  
Acquired : 07 Jul 2021 17:23 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-296S  
Misc Info : Cyromazine in MeOH  
Vial Number: 19



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

