



```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    6
Acq. Instrument : LCMS-1                      Location  : P1-D1
Injection Date  : 3/22/2024 11:10:37 AM      Inj       :    1
                                                Inj Volume: 1.000 µl
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 3.000 µl

```
Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 3/26/2024 2:32:15 PM by SYSTEM
                 (modified after loading)
```

```
Method Info    : Operator: Organic QC lab
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                 /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
                 Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
                 Reference.
```

Sample Info : DCPA Diacid in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

