

=====
Acq. Operator : SYSTEM Seq. Line : 26
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-B-01
Injection Date : 4/19/2021 3:34:20 PM Inj : 1
Inj Volume : 1.000 µl



Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\04192101\04192101 2021-04-19 07-40-22
\Purity_195-315.M
Last changed : 5/8/2020 10:12:19 AM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed : 4/19/2021 4:25:16 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

