

Print of window 38: Current Chromatogram(s)

Data File : D:\DATA\06081801\06081801 2018-06-08 06-50-10\13-Sample Name : P-369S-CN-10X (Thidiazuron)



=====

Acq. Operator : SYSTEM	Seq. Line : 13
Acq. Instrument : LCMS-1	Location : P1-F6
Injection Date : 6/8/2018 10:35:06 AM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : D:\DATA\06081801\06081801 2018-06-08 06-50-10\Purity195_315.M
Last changed : 6/8/2018 6:50:10 AM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 6/11/2018 10:47:07 AM by SYSTEM
(modified after loading)

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water /100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min . Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

