



Acq. Operator : SYSTEM

Seq. Line :

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P1-E2

Injection Date : 4/12/2023 2:18:57 PM

Inj : 1

Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\041223\032123101 2023-04-12 07-41-23\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Dimethyl Phosphate in MeOH

Current Chromatogram(s)

