

```

=====
Acq. Operator   : SYSTEM                               Seq. Line :    2
Acq. Instrument : LCMS-1                             Location  : P1-A2
Injection Date  : 11/10/2022 7:23:54 AM              Inj       :    2
                                                    Inj Volume: 1.000 µl
  
```

```

Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 11/9/2022 8:35:37 AM by SYSTEM
  
```

```

Method Info    : Operator: Organic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                Reference.
  
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

