

Sample Name : P-486S



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    6
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  : P1-E-02
Injection Date  : 6/24/2020 6:18:39 PM        Inj       :    1
                                           Inj Volume: 1.000 µl
  
```

```

Acq. Method    : C:\Users\Public\Documents\ChemStation\1\Data\06242001\06242001 2020-06-24 16-43-02
                \Purity_195-315.M
  
```

```

Last changed   : 5/8/2020 10:12:19 AM by SYSTEM
  
```

```

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
  
```

```

Last changed   : 6/24/2020 12:11:23 PM by SYSTEM
                (modified after loading)
  
```

```

Method Info    : OPERATOR: Organic QC lab
  
```

```

                General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
                Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
                plus C-18, 2.1mmx50mm, 1.8 µ particlue size. Agilent 1290 Infinity II LC/MS system
                , Agilent 6125 quadrupole MS detector.
  
```

```

Additional Info : Peak(s) manually integrated
  
```

Current Chromatogram(s)

