

=====

Acq. Operator	: SYSTEM	Seq. Line	: 6
Acq. Instrument	: LCMS-1	Location	: P1-E8
Injection Date	: 9/1/2022 7:30:01 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195\_315.M

Last changed : 11/5/2021 2:37:17 PM by SYSTEM

Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M

Last changed : 8/31/2022 10:35:03 AM by SYSTEM

(modified after loading)

Method Info

: Operator: Organic QC lab  
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

DAD1 A, Sig=255,120 Ref=400,100 (D:\DATA\09012022\06- -KD.D)

