


File :D:\MassHunter\GCMS\1\DATA\101420\03- KD.D
Operator : Organic QC Lab
Acquired : 14 Oct 2020 06:30 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: P-548S-CN
Misc Info :
Vial Number: 11

 **AccuStandard**[®]
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 μ m
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1 μ L inj.; GC/MS; INJ 270c

