

File : Q:\GCMS-07\DATA\06061801\48- : K.D
Operator
Acquired : 07 Jun 2018 00:25 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: P-575S-10X
Misc Info : Dinitramine @1000ug/mL in MeOH
Vial Number: 67

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

