

=====

Acq. Operator	: SYSTEM	Seq. Line	: 2
Acq. Instrument	: LCMS-1	Location	: P1-A1
Injection Date	: 12/6/2021 4:38:46 PM	Inj	: 2
		Inj Volume	: 1.000 µl
Acq. Method	: C:\CHEM32\1\METHODS\Purity195_315.M		
Last changed	: 11/5/2021 2:37:17 PM by SYSTEM		
Analysis Method	: C:\CHEM32\1\METHODS\Purity195_315.M		
Last changed	: 12/8/2021 4:14:21 PM by SYSTEM		
	(modified after loading)		
Method Info	: Operator: Organic QC lab		
	Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %		
	Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate		
	:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron		
	particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole		
	MS Detector. No UV Reference.		



Sample Info : Guazatine acetate @ 100 ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

