



=====
Acq. Operator : SYSTEM Seq. Line : 3
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-F-02
Injection Date : 6/15/2021 12:28:14 PM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06152101\06152101 2021-06-15 11-49-24
\Purity_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

