

File :Q:\GCMS-08 Minimal\DATA\20210330C\59- YX.D
Operator : Organic QC Lab
Acquired : 31 Mar 2021 22:25 using AcqMethod Hy_Pest_15cm_S2_3ulinj.M
Instrument : GCMS 8
Sample Name: P-718S
Misc Info : Mephosfolan in MeOH
Vial Number: 21



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

