

=====

Acq. Operator	: SYSTEM	Seq. Line	: 8
Acq. Instrument	: LCMS-1	Location	: P1-D9
Injection Date	: 8/25/2022 11:51:31 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method	: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed	: 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method	: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed	: 8/24/2022 10:35:10 AM by SYSTEM

Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Sample Info : Dimethenamid in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

