

Sample Name : P-832S-CN



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :   8
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                     Location  : P1-E-08
Injection Date  : 6/28/2021 9:32:45 AM       Inj       :   1
                                           Inj Volume: 1.000 µl

```

```

Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\06282101\06252101 2021-06-28 07-18-59
                  \Purity_195-315.M
Last changed    : 4/29/2021 10:41:07 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed    : 4/29/2021 10:41:07 AM by SYSTEM
Method Info     : OPERATOR: Organic QC lab
                  General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%
                  Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
                  plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system
                  , Agilent 6125 quadrupole MS detector.

```

Additional Info : Peak(s) manually integrated

## Current Chromatogram(s)

