



=====

Acq. Operator	: SYSTEM	Seq. Line	: 7
Acq. Instrument	: LCMS-1	Location	: P1-A7
Injection Date	: 6/5/2019 3:26:46 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : D:\DATA\06051901\06051901 2019-06-05 13-34-24\Purity195_315.M

Last changed : 6/5/2019 1:34:24 PM by SYSTEM

Analysis Method : C:\Chem32\1\Methods\Purity195_315.M

Last changed : 4/10/2019 10:56:50 AM by SYSTEM

Method Info : Operator: GA
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Spinosad

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

