

File :D:\MassHunter\GCMS\1\data\05031901\13- D  
Operator : KD  
Acquired : 03 May 2019 10:54 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-868S-CN  
Misc Info : Triticonazole in ACCN  
Vial Number: 57

Column: DB-5, 30m, 0.25 ID, 0.25um  
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.  
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

