

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-06 Minimal\DATA\031422\1  
... 4- KM.D  
Operator : Organic QC Lab  
Instrument : GCMS 6  
Acquired : 14 Mar 2022 16:14 using AcqMethod CHICK\_2019\_13MIN.M  
Sample Name: P-874S  
Misc Info : Metalaxyl-M in MeOH



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

