Print of window 38: Current Chromatogram(s)
Data File : C:\USERS\P...TION\1\DATA\12012001\10152001 2020-12-01 14-31-55\09--GA.D
Sample Name : P-902S-CN

==========

Acq. Operator : SYSTEM Seq. Line : 9
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P1-A-09
Injection Date : 12/1/2020 5:04:08 PM Inj : 1
Inj Volume : 1.000 µl
Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\12012001\10152001 2020-12-01 14-31-55 \Purity_195-315.M
Last changed : 5/8/2020 10:12:19 AM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed : 12/2/2020 9:21:23 AM by SYSTEM
(modified after loading)
Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min,the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)