



=====

Acq. Operator	: SYSTEM	Seq. Line	: 2
Acq. Instrument	: LCMS-1	Location	: P1-C5
Injection Date	: 3/26/2024 7:38:31 AM	Inj	: 2
		Inj Volume	: 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

Acq. Method	: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed	: 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method	: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed	: 3/22/2024 10:44:07 AM by SYSTEM

(modified after loading)

Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple MS Detector. No UV Reference.

Sample Info : Alachlor ESA Sodium Salt in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

