



=====
Acq. Operator : SYSTEM Seq. Line : 2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-A-02
Injection Date : 5/27/2022 9:01:35 AM Inj : 2
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\05272201\05262201 2022-05-27
08-41-47\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 5/16/2022 12:35:31 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity
II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Acetochlor ESA Sodium Salt in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

