



Acq. Operator : SYSTEM Seq. Line : 5
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P1-B-03
Injection Date : 10/20/2021 7:18:25 AM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\10202101\10202101 2021-10-20
06-01-57\Purity_195-315.M
Last changed : 4/29/2021 10:41:07 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed : 4/29/2021 10:41:07 AM by SYSTEM
Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity
II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

