

Print of window38: Current Chromatogram(s)

Data File : D:\DATA\06181801\060181801 2018-06-18 16-04-05\ Sample Name :  
P-944S-CN-10XCyhalofop-butyl



=====

Acq. Operator	: SYSTEM	Seq. Line	: 5
Acq. Instrument	: LCMS-1	Location	: P2-A5
Injection Date	: 6/18/2018 5:19:30 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : D:\DATA\06181801\060181801 2018-06-18 16-04-05\Purity195\_315.M  
Last changed : 6/18/2018 4:04:06 PM by SYSTEM  
Analysis Method : C:\Chem32\1\Methods\Purity195\_315.M  
Last changed : 6/18/2018 9:35:46 AM by SYSTEM  
(modified after loading)

Method Info : Operator: GA  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.  
3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule  
size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector.  
No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

