

=====
Acq. Operator : SYSTEM Seq. Line : 2
Acq. Instrument : LCMS-1 Location : P2-B2
Injection Date : 4/1/2020 5:47:19 AM Inj : 2
 Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 10/18/2019 4:18:14 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 3/31/2020 11:02:42 AM by SYSTEM
(modified after loading)

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
particulate size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple
MS Detector. No UV Reference.

Sample Info : Prothioconazole in ACCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

