

```

=====
Acq. Operator   : SYSTEM                               Seq. Line :    3
Acq. Instrument : LCMS-1                             Location  : P1-E2
Injection Date  : 7/22/2024 9:32:53 AM              Inj       :    1
                                                    Inj Volume: 1.000 µl
  
```

```

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method  : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 7/12/2024 8:26:04 AM by SYSTEM
  
```

```

Method Info     : Operator: Organic QC lab
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                  MS Detector. No UV Reference.
  
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

