



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :   10
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  :   P1-E-11
Injection Date  : 2/25/2021 11:58:01 AM              Inj       :    1
                                                    Inj Volume: 2.000 µl
    
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.020 µl

```

Acq. Method    : C:\Users\Public\Documents\ChemStation\1\Data\02252101\0210202101 2021-02-25 09-06-58\PFAM.M
Last changed   : 1/26/2021 10:30:23 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed   : 2/25/2021 9:22:05 AM by SYSTEM
                (modified after loading)
Method Info    : OPERATOR: Organic QC lab
    
```

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

