



```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  : P1-C-05
Injection Date  : 11/17/2021 6:50:51 AM                Inj       :    2
                                                    Inj Volume: 2.000 µl
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.010 µl

```
Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\11172101\11162101 2021-11-17 06-30-59
                  \PFAA.M
```

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

```
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
```

Last changed : 11/16/2021 8:12:29 AM by SYSTEM  
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

