



```
=====
Acq. Operator   : SYSTEM                       Seq. Line :    2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                       Location  :   P1-C2
Injection Date  : 11/3/2021 6:22:33 AM         Inj       :    2
                                           Inj Volume: 2.000 µl
```

```
Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\11032101\11032101 2021-11-03
              06-02-43\PF_AA.M
Last changed : 1/26/2021 10:30:23 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed   : 4/29/2021 10:41:07 AM by SYSTEM
Method Info : OPERATOR: Organic QC lab
              General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
              Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
              Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity
              II LC/MS system, Agilent 6125 quadrupole MS detector.
```

Current Chromatogram(s)

