



Acq. Operator : SYSTEM Seq. Line : 9  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2 Location : P2-A3  
Injection Date : 1/3/2024 12:42:16 PM Inj : 1

Inj Volume : 2.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\010324\010324 2024-01-03 10-10-17  
\PFAA.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 1/2/2024 7:39:28 AM by SYSTEM  
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system  
, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

